



Batch Number: 51917
 Product Type: CBDPet 100
 Unit Size: 60 ml

Testing Lab: Pixis Labs
 Accreditation: Oregon Accreditation: OR100028

LEGEND

NR: Not Reported
 NA: Not Available
 LOQ: Limited of Quantification
 LOD: Limit of Detection
 1 g = 1000 mg = 10⁶ µg

POTENCY	PER UNIT	PER GRAM	REGULATORY ACTION LEVEL
CBD	100 mg	1.81 mg/g	Not Applicable
THC (Total)	2.59 mg	.05 mg/g	3 mg/g

HEAVY METALS	PER UNIT	REGULATORY ACTION LEVEL
Arsenic	< LOQ	1.5 mg/kg
Cadmium	< LOQ	0.5 mg/kg
Lead	< LOQ	0.5 mg/kg
Mercury	< LOQ	3.0 mg/kg

RESIDUAL SOLVENTS	REGULATORY ACTION LEVEL
All analytes Passing and less than LOQ.	Varies - See Page 3

PESTICIDES	REGULATORY ACTION LEVEL
None of the 373 pesticides tested found above LOQ in the sample.	Varies - See Page 5

MICROBIOLOGY	RESULT	STATUS
Aerobic Plate Count	10 cfu/g	TESTED
E.coli	< LOQ	PASS
Total Coliforms	< LOQ	PASS
Staphylococcus aureus	< LOQ	PASS
Mold (RAPID Petrifilm)	< LOQ	PASS
Yeast (RAPID Petrifilm)	< LOQ	PASS
Salmonella spp.	Negative	PASS
Listeria spp.	Negative	PASS
E.coli O157:H7	Negative	PASS



This is an amended version of report# 19-014624/D01.R00.
Reason: Report updated to include batch ID.

Customer: Nutra Pure LLC
Product identity: CBDPet 100
Client/Metric ID: Batch: 51917
Laboratory ID: 19-014624-0001

Summary

Potency:

Analyte per 60ml	Result	Limits	Units	Status	
CBD per 60ml	100		mg/60ml		CBD-Total per 60ml 102 mg/60ml
Δ9-THC per 60ml	2.59		mg/60ml		THC-Total per 60ml < 3.456 mg/60ml
(Reported in milligrams per serving)					

Residual Solvents:

All analytes passing and less than LOQ.

Pesticides:

Analyte	Result (mg/kg)	Limits (mg/kg)	Status
Multi-Residue Pesticide Profile ¹	< LOQ for all analytes		

Metals:

Less than LOQ for all analytes.

Microbiology:

Analyte	Result	Limits
Aerobic Plate Count	10	



Customer: Nutra Pure LLC
Product identity: CBDPet 100
Client/Metric ID: Batch: 51917
Sample Date:
Laboratory ID: 19-014624-0001
Relinquished by: Terry Montgomery
Temp: 14.9 °C
Serving Size #1: 55.2 g

Sample Results

Potency per 60ml		Batch: 1911071					
Analyte	Result	Limits	Units	LOQ	Analyze	Method	Notes
CBC per 60ml [†]	< LOQ		mg/60ml	1.84	12/10/19	J AOAC 2015 V98-6	
CBC-A per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
CBC-Total per 60ml [†]	< LOQ		mg/60ml	3.46	12/10/19	J AOAC 2015 V98-6	
CBD per 60ml	100		mg/60ml	1.84	12/10/19	J AOAC 2015 V98-6	
CBD-A per 60ml	< LOQ		mg/60ml	1.84	12/10/19	J AOAC 2015 V98-6	
CBD-Total per 60ml	102		mg/60ml	3.46	12/10/19	J AOAC 2015 V98-6	
CBDV per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
CBDV-A per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
CBDV-Total per 60ml [†]	< LOQ		mg/60ml	3.43	12/09/19	J AOAC 2015 V98-6	
CBG per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
CBG-A per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
CBG-Total per 60ml [†]	< LOQ		mg/60ml	3.46	12/09/19	J AOAC 2015 V98-6	
CBL per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
CBN per 60ml	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
Δ8-THC per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
Δ9-THC per 60ml	2.59		mg/60ml	1.84	12/10/19	J AOAC 2015 V98-6	
THC-A per 60ml	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
THC-Total per 60ml	< LOQ		mg/60ml	3.46	12/10/19	J AOAC 2015 V98-6	
THCV per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
THCV-A per 60ml [†]	< LOQ		mg/60ml	1.84	12/09/19	J AOAC 2015 V98-6	
THCV-Total per 60ml [†]	< LOQ		mg/60ml	3.43	12/09/19	J AOAC 2015 V98-6	



Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Notes
Aerobic Plate Count	10		cfu/g	10	1910984	12/06/19	AOAC 990.12 (Petrifilm)	
E.coli	< LOQ		cfu/g	10	1910980	12/06/19	AOAC 991.14 (Petrifilm)	
Total Coliforms	< LOQ		cfu/g	10	1910980	12/06/19	AOAC 991.14 (Petrifilm)	
Staphylococcus aureus	< LOQ		cfu/g	10	1910985	12/05/19	AOAC 2003.07	
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	1910983	12/06/19	AOAC 2014.05 (RAPID)	
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	1910983	12/06/19	AOAC 2014.05 (RAPID)	
Salmonella spp.	Negative		/10g		1910988	12/05/19	AOAC 2016.01	
Listeria spp.	Negative		/10g		1910986	12/05/19	AOAC 2016.07	
E.coli O157:H7	Negative		/5g		1910987	12/05/19	AOAC 2017.01	

Solvents		Method EPA5021A				Units µg/g	Batch 1910951	Analyze 12/03/19 12:06 PM			
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethyl acetate	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether	< LOQ	5000	200	pass	
Ethylene glycol	< LOQ	620	200	pass		Ethylene oxide	< LOQ	50.0	30.0	pass	
Hexanes (sum)	< LOQ	290	150	pass		Isopropyl acetate	< LOQ	5000	200	pass	
Isopropylbenzene	< LOQ	70.0	30.0	pass		m,p-Xylene	< LOQ		200		
Methanol	< LOQ	3000	200	pass		Methylene chloride	< LOQ	600	200	pass	
Methylpropane	< LOQ		200			n-Butane	< LOQ		200		
n-Heptane	< LOQ	5000	200	pass		n-Hexane	< LOQ		30.0		
n-Pentane	< LOQ		200			o-Xylene	< LOQ		200		
Pentanes (sum)	< LOQ	5000	600	pass		Propane	< LOQ	5000	200	pass	
Tetrahydrofuran	< LOQ	720	100	pass		Toluene	< LOQ	890	100	pass	
Total Xylenes	< LOQ		400			Total Xylenes and Ethyl	< LOQ	2170	600	pass	

Pesticides		Method AOAC 2007.01 & EN 15662 (mod)			Units mg/kg	Batch 1911039	Analyze 12/05/19 08:20 AM			
Analyte	Result	Limits	Status	Notes						
Multi-Residue Pesticide Profile ¹	< LOQ for all analytes									

Metals

Analyte	Result	Limits	Units	LOQ	Batch	Analyze	Method	Notes
Arsenic	< LOQ		mg/kg	0.0414	1911076	12/05/19	AOAC 2013.06 (mod.)	
Cadmium	< LOQ		mg/kg	0.0414	1911076	12/05/19	AOAC 2013.06 (mod.)	
Lead	< LOQ		mg/kg	0.0414	1911076	12/05/19	AOAC 2013.06 (mod.)	
Mercury	< LOQ		mg/kg	0.0207	1911076	12/05/19	AOAC 2013.06 (mod.)	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Pixis quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.



These test results are representative of the individual sample selected and submitted by the client.

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

† = Analyte not NELAP accredited.

Units of Measure

cfu/g = Colony forming units per gram

g = Gram

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/55.2g = Milligram per 55.2g

/10g = Per 10 grams

/5g = Per 5 grams

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner
General Manager



Fixis Labs
Canrabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)	Compound	LOQ (mg/kg)
Abamectin	0.100	CIPC	1.000	Endrin	0.100
Acephate	0.100	Cethodim	0.050	EPN	0.050
Acequinocyl	0.100	Cethodim Sulfone	0.050	EPTC	0.100
Acetamiprid	0.020	Cethodim Sulfoxide	0.050	Esfenvalerate/ Fenvalerate	0.200
Acetochlor	0.100	Cbfentazine	0.020	Etaconazole	0.100
Acrinathrin	0.100	Cbmazone	0.020	Ethalfuralin	0.100
Alachlor	0.100	Cbthianidin	0.200	Ethiofencarb	0.050
Aldicarb	0.100	Cumaphos	0.050	Ethion	0.200
Aldicarb sulfoxide	0.100	Crdoxyphos	0.020	Ethirimol	0.100
Aldoxycarb (Aldicarb-sulfone)	0.100	Cyarazine	0.020	Ethofumesate	0.050
Aldrin	0.100	Cyarfenphos	0.020	Ethoprophos	0.020
Ametoctradin	0.020	Cyartaniliprole	0.050	Etofenprox	0.020
Ametryn	0.500	Cyazfamid	0.020	Etoxazole	0.020
Aspon	0.100	Cydoate	0.100	Etridiazole	0.100
Asulam	0.100	Cyfluthrin	0.200	Etrimefos	0.020
Atrazine	0.100	Cyhalothrin, lambda	0.200	Famoxadone	0.200
Atrazine-desethyl	0.100	Cymoxanil	0.050	Famphur	0.100
Azinphos-ethyl	0.020	Cypermethrin	0.200	Fenamidone	0.020
Azinphos-methyl	0.020	Cyprodinil	0.100	Fenamiphos	0.020
Azoxystrobin	0.020	Dathal	0.100	Fenamiphos sulfone	0.020
Beralaxyl	0.020	Damnozide	0.100	Fenamiphos sulfoxide	0.020
Berthocarb	0.020	DCPMU	0.050	Fenazaquin	0.100
Berfluralin	0.100	DDD, qp'	0.100	Fenbuconazole	0.100
Berxacor	0.050	DDD, p,p'	0.100	Fenchlorphos	0.100
Bersulide	0.050	DCE, o,p'	0.100	Fenchlorphos-oxon	0.100
BHC alpha isomer	0.100	DCE, p,p'	0.100	Fenhexamid	0.100
BHC beta isomer	0.100	DDT, o,p'	0.100	Fenitrothion	0.100
BHC delta isomer	0.500	DDT, p,p'	0.100	Fenobucarb	0.050
Bifenazate	0.020	DEF (Tribufos)	0.100	Fenoxycarb	0.020
Bifenthrin	0.020	Deltamethrin	0.100	Fenpropathrin	0.050
Boscalid	0.020	Desmedipham	0.100	Fenpyroximate	0.020
Bromophos-ethyl	0.100	Diallate	0.100	Fenson	0.100
Bromophos-methyl	0.200	Diazinon	0.020	Fensulfthion	0.020
Bromopropylate	0.100	Diazoxon	0.100	Fensulfthion oxon	0.020
Bromuconazole	0.100	Dichlobenil	0.100	Fensulfthion sulfone	0.100
Bupirimate	0.020	Dichlofluanid	0.100	Fensulfthion-oxon-sulfone	0.020
Buprofezin	0.050	Dichlorvos	0.100	Fenthion	0.050
Butachlor	0.500	Diclobutrazol	0.050	Fenthion oxon	0.020
Butralin	0.200	Dicofol	0.100	Fenthion oxon sulfone	0.100
Butylate	0.100	Dicrotophos	0.050	Fenthion oxon sulfoxide	0.020
Cadusafos	0.020	Dieldrin	0.100	Fenthion sulfoxide	0.100
Captaf	1.000	Diethofencarb	0.020	Fenthion sulfone	0.050
Cartaryl	0.050	Diethyltoluamide (DEET)	0.050	Fenuron	0.020
Cartendazim	0.100	Difenoconazole	0.100	Fipronil	0.100
Cartofuran	0.020	Dimethenamid	0.050	Fonicamid	0.100
Cartophenothion	0.100	Dimethoate	0.050	Rucloralin	0.100
Cartoxin	0.020	Dimethomorph	0.020	Rucluthrinat	0.100
Carfentrazone-ethyl	0.100	Diniconazole	0.200	Rudioxonil	0.200
Chlorantraniliprole	0.020	Dinotefuran	0.200	Rufenacet	0.020
Chordane, cis	0.200	Dioxathion	0.100	Rumioxazin	0.100
Chordane, trans	0.200	Diphenamid	0.020	Fuometuron	0.020
Chorfenapyr	0.500	Diphenylamine	0.100	Fuopicolide	0.050
Chorfenoson	0.200	Disulfoton	0.100	Fuopyram	0.020
Chorfenvinphos	0.050	Disulfoton sulfone	0.100	Fuoxastrobin	0.050
Chorobenzilate	0.100	Disulfoton sulfoxide	0.100	Fuopyradfurone	0.020
Choroneb	0.200	Diuron	0.050	Furidone	0.100
Chorpyrifos	0.050	Edifenphos	0.050	Fusilazole	0.020
Chorpyrifos-methyl	0.200	Endosulfan alpha	0.200	Flutolanil	0.020
		Endosulfan beta	0.200	Flutriafol	0.020
		Endosulfan sulfate	0.100	Fluvalinate, tau-	0.100
				Fluxapyroxad	0.020



Fixis Labs
Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ(mg/ kg)	Compound	LOQ(mg/ kg)	Compound	LOQ(mg/ kg)
Fomesafen	0.100	Mexacarbate	0.020	Propamocarb	0.050
Fonofos	0.100	MGK 264	0.020	Proparil	0.050
Forchlorfenuron	0.050	Mirex	0.100	Propargite	0.050
Formetanate	0.050	Molinate	0.050	Propazine	0.020
Furathiocarb	0.020	Monocrotophos	0.100	Propetamphos	0.050
Heptachlor	0.100	Monolinuron	0.020	Prorham	0.050
Heptachlor epoxide	0.100	Myclobutanil	0.050	Propiconazole	0.050
Heptenophos	0.100	Naled	0.100	Propoxur	0.050
Hexachlorobenzene	0.100	Napropamide	0.050	Propoxycarbazone Na	0.050
Hexaconazole	0.100	Neburon	0.020	Propyzamide	0.050
Hexazinone	0.100	Nitrapyrin	0.100	Prothiofos	0.100
Hexythiazox	0.020	Norflurazon	0.050	Pyraclostrobin	0.020
Imazalil	0.100	Omethoate	0.100	Pyrazophos	0.050
Imidacloprid	0.100	O-Phenylphenol	0.100	Pyrethrins	0.050
Indaziflam	0.020	Oxadixyl	0.100	Pyridaben	0.020
Indoxacarb	0.020	Oxamyl	0.100	Pyridafol	0.100
Iprobenfos	0.100	Oxamyl-oxime	0.100	Pyridate	0.020
Iprodione	0.100	Oxychlorane	0.100	Pyrimethanil	0.050
Isobenzan	0.100	Oxydemeton-Methyl	0.100	Pyriproxifen	0.020
Isocarbophos	0.500	Oxythioquinox	0.200	Pyrosulfone	0.020
Isodrin	0.100	Paclobutrazol	0.050	Pyroxulam	0.020
Isufenphos	0.050	Paraoxon-ethyl	0.020	Quinalphos	0.050
Isufenphos-methyl	0.020	Paraoxon methyl	0.100	Quinoxifen	0.050
Isufenphos oxon	0.050	Parathion ethyl	0.100	Quintozene (PQNB)	0.200
Isoprocarb	0.020	Parathion methyl	0.200	Resmethrin	0.050
Isopropalin	0.200	Perconazole	0.050	Rotenone	0.050
Isoprothiolane	0.050	Perdimethalin	0.050	S421	0.100
Isoproturon	0.050	Perflufen	0.020	Smazine	0.100
Isoxaben	0.050	Pertachloroaniline	0.100	Smetryn	0.200
Isoxaflutole	0.050	Pertachloroanisole	0.100	Spinetoram	0.020
Kresoxim-methyl	0.050	Pentachlorobenzene (PCB)	0.100	Spinosad	0.050
Ladofen	0.500	Pentachloroanisole (PCTA)	0.100	Spirodiclofen	0.100
Lenadi	0.100	Perthiopyrad	0.020	Spiromesifen	0.050
Lindane (gammaBHC)	0.100	Permethrin	0.050	Spirotetramat	0.050
Linuron	0.020	Pethane	0.100	Spiroxamine	0.020
Malaaxon	0.050	Phenmedipham	0.050	Sulfotep	0.050
Malathion	0.050	Phenthoate	0.050	Sulfoxalfor	0.050
Mandipropamid	0.020	Phorate	0.050	Sulprofos	0.020
Mecarbam	0.020	Phorate Sulfone	0.050	Tebuconazole	0.100
Mepanipyrim	0.050	Phorate Sulfoxide	0.050	Tebufenozide	0.020
Merphos	0.500	Phosalone	0.050	Tebuthiuron	0.020
Metalaxyl	0.050	Phosmet	0.100	Tecnazene	0.100
Metaldehyde	0.050	Phosphamidon	0.050	Tefluthrin	0.100
Metconazole	0.100	Phoxim	0.050	Terbufos	0.020
Methacrifos	0.100	Pinoxaden	0.020	Terbufos sulfone	0.050
Methamidophos	0.050	Piperonyl butoxide	0.050	Terbufos sulfoxide	0.050
Methidathion	0.050	Pirimicarb	0.020	Terbutylazine	0.020
Methiocarb	0.050	Pirimiphos-methyl	0.050	Terbutryn	0.020
Methiocarb sulfone	0.100	Pirimiphos-ethyl	0.020	Tetrachlorvinphos	0.050
Methiocarb sulfoxide	0.100	Prallethrin	0.100	Tetraconazole	0.050
Methomyl	0.100	Prochloraz	0.020	Tetradifon	0.200
Methoxychlor	0.100	Procymidone	0.100	Tetramethrin	0.050
Methoxyfenozide	0.020	Profenofos	0.100	Tetrasul	0.100
Metobromuron	0.050	Prfluralin	0.100	Thiabendazole	0.100
Metolachlor	0.100	Promecarb	0.050	Thiabendazole, 5-hydroxy	0.100
Metolcarb	0.050	Prometon	0.100	Thiadoprid	0.050
Metraferone	0.050	Prometryn	0.020	Thiamethoxam	0.100
Metribuzin	0.100	Propachlor	0.020	Thiobencarb	0.050
Mevinphos	0.100			Thiodicarb	0.050
				Thiophanate-methyl	0.050



PIXIS Labs
Cannabis Multi-Residue Profile, Limits of Quantitation

Compound	LOQ(mg/kg)	Compound	LOQ(mg/kg)	Compound	LOQ(mg/kg)
Tolclofos-methyl	0.100	Triazophos	0.020	Trifloxystrobin	0.020
Triforin	0.100	Tolyfluanid	0.050	Triticonazole	0.050
Tralkoxydim	0.100	Tridiphane	0.500	Vindozolin	0.100
Triadimefon	0.050	Triflumizde	0.020	Zoxamide	0.020
Triallate	0.100	Trifluralin	0.100		

LOQ=Limit of Quantitation, mg/kg

Factors affecting the LOQ include instrumentation sensitivity for a particular analyte, sample size, moisture content (percent solids) of the sample, effectiveness of the cleanup on the sample extract, and especially the type of sample matrix.